

Cementation Guide

Dentist:
Patient:
☐ This is the restoration you are receiving and the respective pre-treatment (large box):

Z Zirconium Oxid
 $\leq 50 \mu\text{m}$ sandblast

Z1 Aluminium oxide 1 bar

Z2 Aluminium oxide 2 bar

ZR 3M™ Rocatec™

Clean, e.g. with air or alcohol

G Glass-Ceramic

Hydrofluoric 5-9,5%

G2 20 sec Hf

G6 60 sec Hf

G9 90 sec Hf

Clean, e.g. rinse with water and air-dry

C Composite
 $\leq 50 \mu\text{m}$ sandblast

Aluminium oxide 1 bar

Clean, e.g. with air or alcohol

M Metal
 $\leq 50 \mu\text{m}$ sandblast

Aluminium oxide 1 bar

Clean, e.g. with air or alcohol

Pre-treatment according to material recommendation already done by the dental technician: ☐ Yes ☐ No

☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐
☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐
18	17	16	15	14	13	12	11	21	22	23	24	25	26	27	28	
48	47	46	45	44	43	42	41	31	32	33	34	35	36	37	38	
☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐
☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐

☐ This is the cementation procedure we recommend (small box):

A Adhesive
☐ 3M™ RelyX™ Universal Resin Cement


See step-by-step guide

(incl. 3M™ Scotchbond™ Universal Plus Adhesive)

☐ 3M™ RelyX™ Veneer Cement


See step-by-step guide

(incl. 3M™ Scotchbond™ Universal Adhesive)

SA Self-adhesive
☐ 3M™ RelyX™ Universal Resin Cement


See step-by-step guide

☐ 3M™ RelyX™ Unicem (2) Cement



See step-by-step guide

C Conventional
☐ 3M™ Ketac™ Cem (Plus) Cement


See step-by-step guide

T Temporary
☐ 3M™ RelyX™ Temp NE Cement


Product information

Available resources


Step-by-step videos for cementation procedures

Please note that the final decision for the appropriate cementation procedure is the responsibility of the treating dental professional. Refer to relevant manufacturers Instructions for Use for complete information.

PLEASE CLEAN AFTER TRY IN ACCORDING TO MATERIAL RECOMMENDATION!

Note:

Date and signature:

